

## **RUBIFIN N°4**

A3 solvent for removal of neat oils, in vacuum or ultrasonic processes.

### MEDICAL RANGE

FUNCTION	APPLICATION/POLLUTION
Vacuum or ultrasonic cleaning	Neat oil

#### COMPATIBILITY

Stainless steel:
Austenitic

– Martensitic

- Ceramic
  - Chromium Cobalt
  - Titanium
- Precipitation hardening

#### COMPONENTS

- A3 solvent
- Hydrocarbon-free
- No CMR compounds, REACH compliant

#### PHYSICOCHEMICAL DATA

- pH concentrated: n.m
- Density: 0.92
- Flash point: 71°C (160°F)
- Surface tension: 29.2 mN/m

#### **INSTRUCTIONS FOR USE\***

- Concentration: 100%
- Temperature: 20 to 50°C (68°F to 122°F)
- Time: 3 to 10 minutes

\*Work in a well ventilated area or one equipped with an aspiration system.

### **PROCESS EXAMPLES**


Removal of oil in vacuum process

CLEANING	RINSING	
RUBIFIN N°4 Vacuum process	RUBIFIN N°4 Vacuum process	DRYING Vacuum process
Conc.: 100% Temp.: 20-50°C 68-122°F Time: 3-10 min	Conc.: 100% Temp.: 20-50°C 68-122°F Time: 3-10 min	
US	US	

Removal of oil in ultrasonic process

#### CLEANING





#### **STORAGE CONDITIONS**

- Keep the recipient hermetically sealed between 5°C and 40°C (41°F and 104°F) in a dry place.
- Always keep in packaging made from the same material as the original packaging (HDPE).

# PRODUCT COMPATIBLE WITH VALIDATION STANDARDS

- Cleaning:
  - ISO 19227 2018
  - ASTM F3127 16



If you have any questions, please contact our Application Centre on: +41 22 365 46 66

NGL CLEANING TECHNOLOGY SA - ECOLOGICAL CLEANING SOLUTIONS Chemin de la Vuarpillière 7 – CH-1260 NYON – SWITZERLAND +41 22 365 46 66 – contact@ngl-group.com – www.ngl-group.com





18/01/23